



Facility: _____ Date Record Started: _____

[illegible]

1. When you begin to use a new meter.
2. When you open a new vial of Glucertain Test Strips.
3. If you think the meter or test strips are not working properly.
4. If the meter is dropped.
5. If test results are higher or lower than expected, and you have repeated the test with similar results.
6. When practicing the test procedure.

Be sure to perform Control Solution Tests on "CS" setting (see Quick Start Guide instruction #1).